

Technical Data Sheet

Doc No – Ab/TDS/Cryo-32

Rev No : 06 Date 02/2022-23

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Cryo Box, PC

Valid For Cat. No. P20601, P20601A, P20602, P20602A, P20603, P20603A

1. Description/Specification

1.1 Description:

1.1.1 Reusable Cryo Box, Made of strong polycarbonate and designed to be used at a temperature between -196°C to 121°C.

1.1.2 Can be used in gas phase of liquid nitrogen with appropriate safety procedure in place.

1.1.3 Transparent lids permit viewing of contents. These boxes create small, space-efficient storage system within a large mechanical freezer.

1.1.4 Lids have numbered grid for sample identification. The surface of the lid has a frosted area suitable for writing with markers, facilitating inventory control.

1.1.5 The dimensions of the cryo box are compatible with vertical, horizontal and drawer style freezer racks.

1.1.6 Specially Designed venting in base & lid ensures a smooth flow for vapours for consistency in cooling.

1.2 Design:

Transparent lid permit viewing design having 25 & 81 Places

1.3 Performance Specifications:

Code	Capacity	Places
P20601	1 & 2 ml	25(5x5)
P20601A	1 & 2 ml	25(5x5)
P20602	1 & 2 ml	81(9x9)
P20602A	1 & 2 ml	81(9x9)
P20603	5 ml	81(9x9)
P20603A	5 ml	81(9x9)



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903

Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslivesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

1.4 Material:

1.4.1 Polycarbonate (PC), confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests, Materials has successfully passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & its following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

1.5 Color:

White, Yellow, & Blue

1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.7 Sterilization:

NA

1.8 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

1.9 Other information

Reusable

2. Basic Features

2.1 Temperature range: -196⁰C to +121⁰C

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 15 minutes

2.3 Manufactured in Aseptic Environment

2.4 Chemical resistance: see website (www.abdoslifesciences.com)



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001, 13485, 14001, 18385, GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakshewari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email: labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 15 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.

Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30⁰C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

WARNING

DO NOT store strong oxidizing agents.

DO NOT place any plastic labware in a flame.

DO NOT mix any chemicals that may result in a thermic reaction, which can cause product failure

Consult local fire codes prior to storage of flammable liquids.

UKRP
Obelis UK LTD
Sandford Gate,
East Point Business Park,
Oxford OX4 8LD
United Kingdom
Tel: +44 1491 378012
info@obelis.co.uk

Obelis s.a.
Bd General Wakis 53
1090 Brussels Belgium
Tel: +(32) 2 732 59 54
Fax: +(32) 2 732 60 03
mail@obelis.net


NON STERILE

30⁰C

EC REG

Abdos Labtech Pvt. Ltd.
Plot No. 1 Lakshwari Bhagwanpur Roorkee 247667 (UK) India
Email : labtech@abdosindia.com, www.abdoslifesciences.com
Ph: +(91) 13 41081897

Electronic IFU



[Signature]
Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakshwari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903
Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7

Technical Data Sheet



Abdos Labtech Private Limited
(A Division of Abdos Group)
(An ISO 9001,13485,14001,18385,GMP & CE certified company)
Shiv Ganga Industrial Estate
Plot No-1, Lakheswari, Bhagwanpur,
Roorkee-247667, (UK) INDIA

Registered office
13A Government
Place East, 2nd Floor Kolkata
West Bengal – 700069
CIN: U85100WB2008PTC123903

Corporate Office
319, 3rd Floor, DLF Tower A, DDA Dist. Centre
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277
Email:labtech@abdosindia.com
Web: www.abdoslifesciences.com
GST: 05AAGCA8605A1Z7