

# Technical Data Sheet

Doc No – Ab/TDS/PCR-114

Rev No: 05 Date 02/2022-23

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**PCR Workstation Rack with Lid, PP**

**Valid for Catalog No. P11103**

### 1. Description/Specification

#### 1.1 Description: PCR Workstation Rack with Lid, PP

1.1.1 These Workstation racks can be easily stackable.

1.1.2 These racks can accommodate unskirted, semi-skirted and skirted 96 well PCR plates, individual PCR tubes and 0.5, 1.5 and 2.0 ml micro centrifuge tube.

1.1.3 These workstations are ideal platform to satisfy all pre or post PCR sampling application.

1.1.4 These Workstations have moulded alphanumeric indexing for fast and accurate identification.

1.1.5 Operating temperature is -80°C to +121°C & can be autoclaved.

#### 1.2 Design:

Rectangular PCR Workstation Rack

#### 1.3 Performance Specification:

Code	Capacity	L x W x H (mm)	No. per Case
<b>P11103</b>	24 x individual 0.2ml PCR tubes, 16 x 0.5,1.5/2.0ml tubes & 96 well rack for PCR plates.	215x120x50	5

#### 1.4 Material:

1.4.1 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests ,Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE ,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

Registered office  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903

Corporate Office  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email: labtech@abdosindia.com  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

## 1.5 Color:

Available in Red, Green, Yellow, White & Blue

## 1.6 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release

## 1.7 Other information

1.7.1 These Workstation racks can be easily stackable.

1.7.2 Recommended for multiple use, reusable.

## 1.8 Basic Features

**1.8.1 Temperature range:** -80 °C to + 121°C

**1.8.2 Autoclavability:** Yes, Autoclavable at 15 psi and 121 °C for 20 minutes

**1.8.3** Manufactured in Aseptic Environment

**1.8.4 Chemical resistance:** see website ([www.abdoslifesciences.com](http://www.abdoslifesciences.com))

## Autoclaving Guideline for Plastics

All plastic consumable are Autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use dry cycle as this may cause deformation or weakening of plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

**Registered office**  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
**Corporate Office**  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email: [labtech@abdosindia.com](mailto:labtech@abdosindia.com)  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

## Instruction for Use

Product can be used for food contact also.  
Storage temp. Max 30°C  
Before using the product check crackness, etc.  
Check all the information printed on pouches or boxes.  
Always wear gloves while handling specimen.

## WARNING

**DO NOT** store strong oxidizing agents.  
**DO NOT** place any plastic labware in a flame.  
**DO NOT** mix any chemicals that may result in a thermic reaction, which can cause product failure  
Consult local fire codes prior to storage of flammable liquids.

**UKRP**  
Obelis UK LTD  
Sandford Gate,  
East Point Business Park,  
Oxford OX4 6LB  
United Kingdom  
Tel: +44 1491 378022  
info@obelis.co.uk

**EC REP**  
Obelis s.a.  
Ed Coneral Wafels 53  
1030 Brussels, Belgium  
Tel: +32 2 732 59 54  
Fax: +32 2 732 60 03  
mail@obelis.net

**NON STERILE**

30°C

**Abdos Labtech Pvt. Ltd.**  
Plot No. 1 Lakshwari Bhagwanpur Roorkee 247667 (UK) India  
Email: labtech@abdosindia.com, www.abdoslifesciences.com  
Ph: +91 31 41081897

Haryana (IN)



Quality Assurance Incharge  
Abdos Labtech Pvt. Ltd.

